

Test Report

No.: QDHL1901001660OT

Date: FEB.12,2019

Page: 1 of 2

SHANDONG INTCO MEDICAL PRODUCTS CO., LTD

NO.9888,QIWANG ROAD,NAOSHAN INDUSTRY PARK,QINGZHOU,SHANDONG,CHINA

The following sample(s) was/were submitted and identified by the client as:

Sample Description : DISPOSABLE VINYL GLOVE (PVC) COLOR: CLEAR
 SGS Ref. No. : KV-19-00637
 Sample Receiving Date : JAN.16,2019
 Testing Period : JAN.16,2019 TO FEB.12,2019
 Test Performed : SELECTED TEST(S) AS REQUESTED BY APPLICANT
 Test Requested : BS EN 455-3-2015 MEDICAL GLOVES FOR SINGLE USE—PART 3:
 REQUIREMENTS AND TESTING FOR BIOLOGICAL EVALUATION
 CLAUSE 4.4 & CLAUSE 4.5
 Test Result(s) : PLEASE REFER TO THE FOLLOWING PAGE(S)
 Conclusion : THE TEST DATA WERE PROVIDED TO CLIENT FOR THEIR OWN
 ANALYSIS.

Remark: Unless otherwise stated the results shown in this test report refer only to the sample(s) tested.

SGS-CSTC Standards Technical Services (Qingdao) Co., Ltd.

Approved by:



Zhou Xinkuan, SK
Lab Manager



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law.
 Attention: To check the authenticity of testing /inspection report & certificate, please contact us at telephone: (86-755)83071443, or email: CN.Doccheck@sgs.com

SGS-CSTC Standards Technical Services (Qingdao) Co., Ltd.

SGS Center, No. 143, Zhuzhou Road, Laoshan District, Qingdao, China 266101
 t (86-532) 68999888 f (86-532) 80991955

www.sgsgroup.com.cn
sgs.china@sgs.com

Test Report

No.: QDHL1901001660OT

Date: FEB.12,2019

Page: 2 of 2

Test Conducted:

BS EN 455-3:2015 Medical gloves for single use—Part 3:Requirements and testing for biological evaluation

Number of test sample	:	5 Pieces
Finishes of gloves	:	Powdered-free gloves other than surgeon's gloves
Defects observed before testing	:	No defects
Test Result	:	Pass

Clause	Test Items	Result	Note
4.4	Powder-free gloves	Pass	# 1
4.5	Proteins, leachable	---	# 2

Notes : #1 Test according to EN ISO 21171:2006, the average mass of powder per glove was 0.02mg.
#2 See result page

Result:

INSPECTION ITEM	TEST RESULT
Aqueous Extractable Protein (ppm)	N.D.

Note: 1. N.D. = Not Detected.

2. MDL (METHOD DETECTION LIMIT):0.2ppm.

Sample Photo:



SGS authenticate the photo on original report only

End of Report



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Document.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law.
Attention: To check the authenticity of testing /inspection report & certificate, please contact us at telephone: (86-755)83071443, or email: CN.Doccheck@sgs.com

SGS-CSTC Standards Technical Services (Qingdao) Co., Ltd.

SGS Center, No. 143, Zhuzhou Road, Laoshan District, Qingdao, China 266101
t (86-532) 68999888 f (86-532) 80991955

www.sgsgroup.com.cn
sgs.china@sgs.com